



REPUBLIKA E SHQIPËRISË
MINISTRIA E SHËNDETËSISË
DHE MBROJTJES SOCIALE

CLARIFICATION NO 1

Issued on 06 July 2023

to the bidding documents for the tender Medical Equipment for Hospitals:

Lot 1 Apheresis

Lot 2 Digital mammography

Lot 3 Ophthalmology

Lot 4 Phaco equipment

No.	Question	Reference to the Bidding Documents /Legal Basis	Answer/Clarification
1.	<p><i>For LOT2:</i></p> <p>It is requested “Motorized height adjustment approx. 75 cm, with electromagnetic brake” Different manufacturer offers different gantry height adjustment range.</p> <p>Do you accept height adjustment range 750mm~1330mm and make competition open?</p>	Bidding Documents/ Technical Specifications	<p><i>Range 750mm ~1330mm includes our request which is “Motorized height adjustment approx. 75 cm,”</i></p>
2	<p><i>For LOT2:</i></p> <p>Adjustment keys on both sides of the compression. Equipment for fine tuning Different manufacturer offers different solution for performing breast compression and fine tuning. Some manufacturer offers compression keys on gantry side and other offers foot-switch pedal for compression. Also pressure sensor and thickness sensor are alternative way for fine tuning and in principle compression + fine tuning are done but in alternative way.</p>	Bidding Documents/ Technical Specifications	<p><i>Based on our market research, there are many manufacturers which fulfill this specification.</i></p>

	Do you accept alternative manufacturer solution for performing breast compression and making competition open?		
3	<p><i>For LOT2:</i></p> <p>Detector section it is requested: Detector with at least 60% DQE measured under 10mR dose.</p> <p>DQE (Detective Quantum Efficiency) measures combined effects of the signal (related to image contrast) and noise performance of an imaging system, generally expressed as a function of spatial frequency measured to lp/mm and not by dose in mR(mili Roengent).</p> <p>Could you please check and correct the requirement for DQE related to lp/mm?</p>	Bidding Documents/ Technical Specifications	<i>The requirement of this specification will be evaluated according to the method used by the manufacturer to assess its DQE of the device that will be offered.</i>
4	<p><i>For LOT2:</i></p> <p>Please clarify if the MAF, CE, ISO are requested only for the main unit mammography and for items such as UPS and printing films are not obligatory?</p>	Bidding Documents/ Technical Specifications	<i>MAF is requested only for Mammography Unit. CE and ISO stands for items such as UPS and printing films.</i>